

TECHNICAL SPECIFICATIONS AND OTHER ALLIED REQUIREMENTS

| SI No. | Description of items | Quantity |
|--------------------------|--|----------|
| PUR/231/ASL/CK/P/2021-22 | | |
| 01. | SUPPLY OF FORCE AND TORQUE SENSOR HAVING THE FOLLOWING SPECIFICATIONS : F/T transducer having the following loads and torques limits Fx, Fy = <20N, with a Resolution of <0.025 Fz=<30N, with a Resolution of <0.05 Tx, Ty, Tz = 100 Nmm, with a Resolution of <0.01 With 1.8 m cable length, Interface board and power supply with cables from power supply or interface-power supply box to DAQ card, unterminated 2m length. | 01 No. |

1. INSTALLATION/COMMISSIONING

1.1. Not required.

2. WARRANTY

2.1. 12 (twelve) months from the date of Acceptance of the Ordered goods.

3. PAYMENT TERMS

3.1. 100% Payment shall be released within 30 days of delivery and acceptance of the ordered goods.

Note: 1. Bid Securing Declaration is to be submitted by the bidder as per format prescribed in the tender documents.