

TECHNICAL SPECIFICATIONS AND OTHER ALLIED REQUIREMENTS

SI No.	Description of items	Quantity
PUR/52/MPML/CK/E/2020-21		
01.	<p>Laser Power Meter</p> <p>Specification:</p> <p>Wavelength range to be measured (μm): 0.25-11 (or Wider)</p> <p>Power range to be measured (W): 100-5,000 (or Wider)</p> <p>Response time (s): <40</p> <p>Resolution of the measured power (W): <10</p> <p>Cooling arrangement: Air/Water cooled</p> <p>Aperture diameter (mm): >30, <60</p>	01 No.

1. INSTALLATION/COMMISSIONING

1.1. NOT APPLICABLE

2. WARRANTY

2.1. One Year from the acceptance of ordered goods.